



Item number 4424

Classic

Examination gloves, Abena Classic, S, transparent, vinyl, powder free

- ✓ Smooth surface
- ✓ No latex proteins or Accelerators



Product Description

Examination glove with smooth surface and contains no latex proteins or accelerators. Suitable for handling non-fatty food products and brief cleaning tasks without strong chemicals.



LOW GRIP



240 mm



EN EN
455 420

EN374-5:2016



VIRUS

374-1:2016/
Type B



KST

AQL
1.5

Specifications

Base name	Examination gloves
Brand	Abena
Sub-brand	Classic
Size	S
Sub-color	Transparent
Properties	Latex free, smooth, low grip, rolled cuff
Features	Powder free
Single or multiple use	Single use
Material	Vinyl
Technical ingredients	PVC, DINP, Viscosity reducer JNJ, PU-coating, Ca-Zn Stabilizer
Length/depth	240 mm
Width	85 mm
Thickness	0,16 mm
Certifications	CE, Food contact materials, CAT III
CE Category (Personal Protective Equipment)	CAT III
CE Class (Medical Devices)	Class I
Product or test standards	EN 455, EN 420, EN 374-1:2016 Type B KST, EN 374-5:2016 VIRUS, AQL 1.5
Directives, regulations and acts	93/42/EEC, 89/686/EEC, 10/2011/EU
Safety Instructions and Warnings	Avoid contact with fatty food products as the plasticizers in the glove can migrate to food during prolonged contact. Contain PVC and plasticizers.
Shelf Life	5 years
Storage Instructions	Store dry, clean and at room temperature.
Product Disposal Instructions	Sort as plastic. If contaminated dispose off as clinical waste.

Instructions for use/application

Examine the gloves for flaws and defects before use

Packaging data

Unit	Contains	Length	Width	Height	EAN
cII	10 pck	31.5 cm	25.5 cm	25 cm	5703538937637
pck	100 pcs	24 cm	12 cm	60 cm	5703538937613
pcs					



The glass fork symbol guarantees that products have been tested in accordance with European legislation and approved for food contact. The symbol is mandatory on products used for food contact.



The CE mark guarantees that a product is safe to use and complies with all safety precautions. CE stands for Conformité Européenne (European Conformity) and is mainly found on electronic equipment, safety equipment, construction products and medical equipment.



Third-party type approval is required for all Category III personal protective equipment (PPE) products. The showing of the CE logo for chemical protective gloves requires that tests are carried out in accordance with test standards specified in EN ISO 374-1: 2016 + A1: 2018 - such as EN 16523-1: 2015 + A1: 2018 to determine the resistance to the permeation of chemicals. The results of this test determine the relevant pictogram symbols that can be used on the packaging and labeling.



The glove has a smooth surface.



The glove has a low grip.



The glove is 240 mm long.



The product does not contain latex.



EN 455 consists of four standards that gloves must be tested against in order to be a medical glove. This standard specifies the requirements for disposable medical gloves.



The standard specifies the requirements for protective gloves in relation to general requirements and test methods.

EN374-5:2016



The standard specifies the requirements for protective gloves against hazardous chemicals and microorganisms. This part of the standard describes the glove's resistance to bacteria, fungi and viruses.



The standard specifies the requirements for protective gloves against chemicals and microorganisms. Type B has been tested for a permeation time of at least 30 minutes with the chemicals sodium hydroxide 40%, hydrochloric acid 40% and formaldehyde 37%.

